


<b>MEK-3DN</b>	HEMATOLOGY CONTROL	2.0mL
----------------	--------------------	-------

**ASSAY SHEET**

LOT	B245N
	2024-08-05
OPEN	14DAYS

<b>MEK-1301,1302,1305 Barcode Entry</b>
---

The barcodes may be read in any order.  
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400	MEK -6500	MEK -1301
			-6410	-6510	-1302
			-6420	-6550	-1305
			-6450		
<b>WBC</b>	10 <sup>3</sup> /μL	±0.8	7.7	7.7	7.71
<b>RBC</b>	10 <sup>6</sup> /μL	±0.20	4.71	4.71	4.72
<b>HGB</b>	g/dL	±0.5	14.0	14.1	14.01
<b>HCT</b>	%	±3.0	40.0	40.0	45.8
<b>MCV</b>	fL	±7.0	84.9	84.9	97.0
<b>MCH</b>	pg	±2.4	29.7	29.9	29.7
<b>MCHC</b>	g/dL	±3.0	35.0	35.3	30.6
<b>PLT</b>	10 <sup>3</sup> /μL	±50	260	260	279.2
<b>MPV</b>	fL	±2.0	5.6	8.1	8.3
<b>RDW-CV *</b>	%	±3.0	16.7	16.7	16.4
<b>LY%</b>	%	±10.0	39.8	39.8	36.13
<b>MO%</b>	%	±5.0	5.0	5.0	1.98
<b>GR%</b>	%	±10.0	55.2	55.2	61.89
<b>LY</b>	10 <sup>3</sup> /μL	±1.0	3.1	3.1	2.79
<b>MO</b>	10 <sup>3</sup> /μL	±0.4	0.4	0.4	0.15
<b>GR</b>	10 <sup>3</sup> /μL	±1.0	4.2	4.2	4.77

<b>MEK-1305</b>
QC ESR
10±5 mm

Manual Registration	
Lot	245
Level	Normal
Control	MEK-3D

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN



MEK-3DN



B245N




2024-08-05



0694-901685P

**MEK-3DH** HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B245H
	2024-08-05
OPEN	14DAYS

**MEK-1301,1302,1305 Barcode Entry**

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 <sup>3</sup> /μL	±2.2	21.0	21.0	20.57
RBC	10 <sup>6</sup> /μL	±0.25	5.95	5.95	5.95
HGB	g/dL	±0.7	19.6	19.5	19.54
HCT	%	±4.0	54.4	54.4	62.1
MCV	fL	±7.0	91.4	91.4	104.4
MCH	pg	±2.8	32.9	32.8	32.8
MCHC	g/dL	±3.0	36.0	35.8	31.5
PLT	10 <sup>3</sup> /μL	±70	543	543	546.9
MPV	fL	±3.0	5.6	8.1	8.5
RDW-CV *	%	±3.0	16.1	16.1	17.6
LY%	%	±10.0	17.3	17.3	13.28
MO%	%	±5.0	2.1	2.1	2.11
GR%	%	±10.0	80.6	80.6	84.61
LY	10 <sup>3</sup> /μL	±1.8	3.6	3.6	2.73
MO	10 <sup>3</sup> /μL	±0.7	0.4	0.4	0.43
GR	10 <sup>3</sup> /μL	±2.0	17.0	17.0	17.41

Manual Registration	
Lot	245
Level	High
Control	MEK-3D

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN



MEK-3DH



B245H




2024-08-05



0694-901686P

**MEK-3DL** HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B245L
	2024-08-05
OPEN	14DAYS

**MEK-1301,1302,1305 Barcode Entry**

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 <sup>3</sup> /μL	±0.6	2.1	2.1	2.05
RBC	10 <sup>6</sup> /μL	±0.15	2.39	2.39	2.32
HGB	g/dL	±0.4	6.0	6.2	6.09
HCT	%	±2.0	17.8	17.8	19.7
MCV	fL	±9.0	74.5	74.5	84.9
MCH	pg	±2.4	25.1	25.9	26.3
MCHC	g/dL	±3.0	33.7	34.8	30.9
PLT	10 <sup>3</sup> /μL	±30	70	70	78.2
MPV	fL	±3.0	6.1	8.4	8.4
RDW-CV *	%	±3.0	17.5	17.5	15.8
LY%	%	±10.0	63.4	63.4	66.96
MO%	%	±5.0	8.4	8.4	4.51
GR%	%	±10.0	28.2	28.2	28.53
LY	10 <sup>3</sup> /μL	±0.5	1.3	1.3	1.37
MO	10 <sup>3</sup> /μL	±0.2	0.2	0.2	0.09
GR	10 <sup>3</sup> /μL	±0.5	0.6	0.6	0.59

<b>MEK-1305</b>
QC ESR
48±17 mm

Manual Registration	
Lot	245
Level	Low
Control	MEK-3D

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN



0694-901684P