


**MK-RE2**
**HEMATOLOGY CONTROL**  
 for reticulocyte

3.0mL

**ASSAY SHEET**

<b>LOT</b>	MKRE247-2
	2024-09-05
<b>OPEN</b>	14 DAYS

		MEK-9200	RANGE
<b>RET%</b>	<b>%</b>	<b>4.00</b>	<b>±3.00</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.1418</b>	<b>±0.1063</b>

**2D Barcode Entry**


• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
**Manual Registration**

Lot	247					
Level	2					
Expiration	9	M	5	D	2024	Y
Control	MKRE					


**CE**  
 0123

**IVD**

**NIHON KOHDEN CORPORATION**  
 1-31-4 NISHIOCHIAI, SHINJUKU-KU  
 TOKYO 161-8560 JAPAN

**MK-RE3 HEMATOLOGY CONTROL**  
for reticulocyte **3.0mL**

**ASSAY SHEET**

<b>LOT</b>	MKRE247-3
	2024-09-05
<b>OPEN</b>	14 DAYS

		<b>MEK-9200</b>	<b>RANGE</b>
<b>RET%</b>	<b>%</b>	<b>7.78</b>	<b>±4.22</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.2807</b>	<b>±0.1522</b>

**2D Barcode Entry**



• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>


<b>Manual Registration</b>						
Lot	247					
Level	3					
Expiration	9	M	5	D	2024	Y
Control	MKRE					



**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN

**MK-RE1 HEMATOLOGY CONTROL**  
for reticulocyte **3.0mL**

**ASSAY SHEET**

<b>LOT</b>	MKRE247-1
	2024-09-05
<b>OPEN</b>	14 DAYS

		<b>MEK-9200</b>	<b>RANGE</b>
<b>RET%</b>	<b>%</b>	<b>1.56</b>	<b>±1.44</b>
<b>RET</b>	<b>10<sup>6</sup>/μL</b>	<b>0.0561</b>	<b>±0.0518</b>

**2D Barcode Entry**



• If RBC value is under 3.00 [ $\times 10^6/\mu\text{L}$ ], please refrain from using.  
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

<b>Manual Registration</b>						
Lot	247					
Level	1					
Expiration	9	M	5	D	2024	Y
Control	MKRE					



 **NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN