


MEK-3DN HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B248N
	2024-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	8.0	8.0	7.90
RBC	10 ⁶ /μL	±0.20	4.83	4.83	4.83
HGB	g/dL	±0.5	14.0	14.2	14.27
HCT	%	±3.0	41.6	41.6	47.6
MCV	fL	±7.0	86.1	86.1	98.5
MCH	pg	±2.4	29.0	29.4	29.5
MCHC	g/dL	±3.0	33.7	34.1	30.0
PLT	10 ³ /μL	±50	249	249	268.7
MPV	fL	±2.0	5.9	8.2	8.5
RDW-CV *	%	±3.0	16.6	16.6	16.4
LY%	%	±10.0	32.9	32.9	30.11
MO%	%	±5.0	3.7	3.7	3.31
GR%	%	±10.0	63.4	63.4	66.58
LY	10 ³ /μL	±1.0	2.6	2.6	2.38
MO	10 ³ /μL	±0.4	0.3	0.3	0.26
GR	10 ³ /μL	±1.0	5.1	5.1	5.26

MEK-1305
QC ESR
8±5 mm

Manual Registration	
Lot	248
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>




NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901685Q

MEK-3DH HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B248H
	2024-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.1	21.1	20.61
RBC	10 ⁶ /μL	±0.25	5.91	5.91	5.94
HGB	g/dL	±0.7	19.2	19.1	19.37
HCT	%	±4.0	55.1	55.1	63.6
MCV	fL	±7.0	93.3	93.3	107.1
MCH	pg	±2.8	32.5	32.3	32.6
MCHC	g/dL	±3.0	34.8	34.7	30.5
PLT	10 ³ /μL	±70	530	530	548.9
MPV	fL	±3.0	5.7	8.1	8.7
RDW-CV *	%	±3.0	15.8	15.8	17.2
LY%	%	±10.0	16.7	16.7	13.18
MO%	%	±5.0	3.4	3.4	2.50
GR%	%	±10.0	79.9	79.9	84.32
LY	10 ³ /μL	±1.8	3.5	3.5	2.72
MO	10 ³ /μL	±0.7	0.7	0.7	0.52
GR	10 ³ /μL	±2.0	16.9	16.9	17.37

Manual Registration	
Lot	248
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>




NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901686Q

MEK-3DL	HEMATOLOGY CONTROL	2.0mL
----------------	--------------------	-------

ASSAY SHEET

LOT	B248L
	2024-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400	MEK -6500	MEK -1301
			-6410	-6510	-1302
			-6420	-6550	-1305
			-6450		
WBC	10 ³ /μL	±0.6	2.0	2.0	2.02
RBC	10 ⁶ /μL	±0.15	2.32	2.32	2.26
HGB	g/dL	±0.4	5.9	6.0	6.03
HCT	%	±2.0	18.1	18.1	20.0
MCV	fL	±9.0	77.9	77.9	88.5
MCH	pg	±2.4	25.4	25.9	26.7
MCHC	g/dL	±3.0	32.6	33.1	30.2
PLT	10 ³ /μL	±30	68	68	75.4
MPV	fL	±3.0	6.6	8.7	8.6
RDW-CV *	%	±3.0	17.2	17.2	15.9
LY%	%	±10.0	64.4	64.4	64.99
MO%	%	±5.0	10.1	10.1	6.37
GR%	%	±10.0	25.5	25.5	28.64
LY	10 ³ /μL	±0.5	1.3	1.3	1.31
MO	10 ³ /μL	±0.2	0.2	0.2	0.13
GR	10 ³ /μL	±0.5	0.5	0.5	0.58

MEK-1305
QC ESR
49±17 mm

Manual Registration	
Lot	248
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN



0694-901684Q