

**MEK-CAL** HEMATOLOGY CALIBRATOR 2.0mL

**ASSAY SHEET**

LOT	PLUS249
	2024-10-05
OPEN	7DAYS

MEK-9100,9200  
2D Barcode Entry



MEK-1301,1302,1303,1305  
Barcode Entry



		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 <sup>3</sup> /μL	9.2	9.2	9.4	9.3	9.1	9.44	9.22
RBC	10 <sup>6</sup> /μL	4.66	4.66	4.62	4.67	4.62	4.60	4.62
HGB	g/dL	13.3	13.4	13.2	13.2	13.1	13.38	13.37
HCT	%	39.7	39.7	41.5	40.3	39.3	44.7	45.0
MCV	fL	85.3	85.3	89.9	86.4	85.1	97.1	97.4
MCH	pg	28.5	28.8	28.6	28.3	28.4	29.1	28.9
MCHC	g/dL	33.5	33.8	31.8	32.8	33.3	29.9	29.7
PLT	10 <sup>3</sup> /μL	238	238	233	236	235	246.9	256.8
MPV	fL	6.8	8.8	6.8	7.6	6.8	9.2	9.2
RDW-CV*	%	16.2	16.2	15.0	16.4	15.3	16.5	16.4

MEK	-7222
	-7300
	-8222
	-9100
	-9200
FS	170
FL	86
SD	114
FS THR.	100

\* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,7222,8222.

<https://www.nihonkohden.com/spi/index.html>

**NIHON KOHDEN CORPORATION**  
1-31-4 NISHIOCHIAI, SHINJUKU-KU  
TOKYO 161-8560 JAPAN

