


MK-RE2
HEMATOLOGY CONTROL
 for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE24Y-2
	2025-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.31	±2.69
RET	10⁶/μL	0.1516	±0.0946

2D Barcode Entry


• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
Manual Registration

Lot	24Y					
Level	2					
Expiration	1	M	5	D	2025	Y
Control	MKRE					

CE
 0123


IVD

NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN


MK-RE3
HEMATOLOGY CONTROL
 for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE24Y-3
	2025-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	8.83	±3.17
RET	10⁶/μL	0.3127	±0.1123

2D Barcode Entry


• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
Manual Registration

Lot	24Y					
Level	3					
Expiration	1	M	5	D	2025	Y
Control	MKRE					

CE
 0123


IVD

NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN

MK-RE1
HEMATOLOGY CONTROL
 for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE24Y-1
	2025-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	1.58	±1.42
RET	10⁶/μL	0.0570	±0.0512

2D Barcode Entry


• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
Manual Registration

Lot	24Y					
Level	1					
Expiration	1	M	5	D	2025	Y
Control	MKRE					


 0123



NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN