


MEK-3DN	HEMATOLOGY CONTROL	2.0mL
----------------	---------------------------	--------------

ASSAY SHEET

LOT	B252N
	2025-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	8.0	8.0	7.88
RBC	10 ⁶ /μL	±0.20	4.90	4.90	4.93
HGB	g/dL	±0.5	14.3	14.5	14.34
HCT	%	±3.0	41.8	41.8	48.0
MCV	fL	±7.0	85.3	85.3	97.4
MCH	pg	±2.4	29.2	29.6	29.1
MCHC	g/dL	±3.0	34.2	34.7	29.9
PLT	10 ³ /μL	±50	252	252	269.1
MPV	fL	±2.0	5.3	8.0	8.2
RDW-CV *	%	±3.0	16.7	16.7	16.7
LY%	%	±10.0	31.8	31.8	28.74
MO%	%	±5.0	3.9	3.9	4.14
GR%	%	±10.0	64.3	64.3	67.12
LY	10 ³ /μL	±1.0	2.5	2.5	2.26
MO	10 ³ /μL	±0.4	0.3	0.3	0.33
GR	10 ³ /μL	±1.0	5.2	5.2	5.29

MEK-1305
QC ESR
7±5 mm

Manual Registration	
Lot	252
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>




NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN



MEK-3DH HEMATOLOGY CONTROL 2.0mL


ASSAY SHEET

LOT	B252H
	2025-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.6	21.6	20.86
RBC	10 ⁶ /μL	±0.25	5.92	5.92	5.91
HGB	g/dL	±0.7	19.3	19.3	19.15
HCT	%	±4.0	54.5	54.5	62.4
MCV	fL	±7.0	92.1	92.1	105.5
MCH	pg	±2.8	32.6	32.6	32.4
MCHC	g/dL	±3.0	35.4	35.4	30.7
PLT	10 ³ /μL	±70	554	554	554.3
MPV	fL	±3.0	5.2	7.9	8.3
RDW-CV *	%	±3.0	16.3	16.3	17.5
LY%	%	±10.0	15.0	15.0	11.30
MO%	%	±5.0	1.5	1.5	1.59
GR%	%	±10.0	83.5	83.5	87.11
LY	10 ³ /μL	±1.8	3.2	3.2	2.36
MO	10 ³ /μL	±0.7	0.3	0.3	0.33
GR	10 ³ /μL	±2.0	18.1	18.1	18.17

Manual Registration	
Lot	252
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>




NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901686S

MEK-3DL	HEMATOLOGY CONTROL	2.0mL
----------------	--------------------	-------

ASSAY SHEET

LOT	B252L
	2025-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.1	2.1	2.08
RBC	10 ⁶ /μL	±0.15	2.47	2.47	2.43
HGB	g/dL	±0.4	6.0	6.2	6.08
HCT	%	±2.0	18.5	18.5	20.6
MCV	fL	±9.0	74.7	74.7	84.9
MCH	pg	±2.4	24.3	25.1	25.0
MCHC	g/dL	±3.0	32.4	33.5	29.5
PLT	10 ³ /μL	±30	66	66	75.6
MPV	fL	±3.0	5.6	8.2	8.2
RDW-CV *	%	±3.0	17.5	17.5	16.0
LY%	%	±10.0	60.8	60.8	59.75
MO%	%	±5.0	7.5	7.5	5.16
GR%	%	±10.0	31.7	31.7	35.09
LY	10 ³ /μL	±0.5	1.3	1.3	1.24
MO	10 ³ /μL	±0.2	0.2	0.2	0.11
GR	10 ³ /μL	±0.5	0.6	0.6	0.73

MEK-1305
QC ESR
43±17 mm

Manual Registration	
Lot	252
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN



0694-901684S