

MEK-CAL HEMATOLOGY CALIBRATOR 2.0mL

ASSAY SHEET

LOT	PLUS255
	2025-06-05
OPEN	7DAYS

MEK-9100,9200
2D Barcode Entry



		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7222	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 ³ /μL	9.3	9.3	9.5	9.4	9.2	9.67	9.26
RBC	10 ⁶ /μL	4.59	4.59	4.58	4.62	4.57	4.55	4.58
HGB	g/dL	13.5	13.6	13.3	13.3	13.2	13.51	13.43
HCT	%	40.0	40.0	42.0	40.7	39.6	45.4	45.5
MCV	fL	87.1	87.1	91.6	88.2	86.7	99.8	99.4
MCH	pg	29.4	29.6	29.0	28.8	28.9	29.7	29.3
MCHC	g/dL	33.8	34.0	31.7	32.7	33.3	29.8	29.5
PLT	10 ³ /μL	226	226	224	227	226	241.8	250.1
MPV	fL	6.0	8.4	6.0	7.3	6.0	8.8	8.9
RDW-CV*	%	15.9	15.9	14.7	15.9	15.0	15.8	15.7

MEK -7222 -7300 -8222 -9100 -9200	
FS	163
FL	90
SD	103
FS THR.	100

MEK-1301,1302,1303,1305
Barcode Entry

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,7222,8222.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-CAL



PLUS255



2025-06-05



0694-901832