

MEK-3DN HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B258N
	2025-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	7.9	7.9	7.95
RBC	10 ⁶ /μL	±0.20	4.78	4.78	4.73
HGB	g/dL	±0.5	13.8	14.0	14.09
HCT	%	±3.0	41.3	41.3	46.7
MCV	fL	±7.0	86.3	86.3	98.8
MCH	pg	±2.4	28.9	29.3	29.8
MCHC	g/dL	±3.0	33.4	33.9	30.2
PLT	10 ³ /μL	±50	242	242	263.1
MPV	fL	±2.0	5.5	8.1	8.2
RDW-CV *	%	±3.0	16.8	16.8	16.3
LY%	%	±10.0	31.9	31.9	31.03
MO%	%	±5.0	2.7	2.7	2.93
GR%	%	±10.0	65.4	65.4	66.04
LY	10 ³ /μL	±1.0	2.5	2.5	2.47
MO	10 ³ /μL	±0.4	0.2	0.2	0.23
GR	10 ³ /μL	±1.0	5.2	5.2	5.25

MEK-1305
QC ESR
8±5 mm

Manual Registration	
Lot	258
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901685U

MEK-3DH	HEMATOLOGY CONTROL	2.0mL
----------------	---------------------------	--------------

ASSAY SHEET

LOT	B258H
	2025-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1


2


3


		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	21.3	21.3	21.06
RBC	10 ⁶ /μL	±0.25	5.90	5.90	5.86
HGB	g/dL	±0.7	18.8	18.8	19.09
HCT	%	±4.0	54.5	54.5	61.9
MCV	fL	±7.0	92.3	92.3	105.7
MCH	pg	±2.8	31.9	31.9	32.6
MCHC	g/dL	±3.0	34.5	34.5	30.8
PLT	10 ³ /μL	±70	519	519	540.8
MPV	fL	±3.0	5.6	8.2	8.5
RDW-CV *	%	±3.0	16.5	16.5	17.6
LY%	%	±10.0	17.1	17.1	14.48
MO%	%	±5.0	1.2	1.2	1.33
GR%	%	±10.0	81.7	81.7	84.19
LY	10 ³ /μL	±1.8	3.6	3.6	3.05
MO	10 ³ /μL	±0.7	0.3	0.3	0.28
GR	10 ³ /μL	±2.0	17.4	17.4	17.73

Manual Registration	
Lot	258
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.

 <https://www.nihonkohden.com/spi/index.html>

 NIHON KOHDEN CORPORATION 1-31-4 NISHIOCHIAI, SHINJUKU-KU TOKYO 161-8560 JAPAN
--



MEK-3DL HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B258L
	2025-11-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.1	2.1	2.04
RBC	10 ⁶ /μL	±0.15	2.33	2.33	2.26
HGB	g/dL	±0.4	5.8	6.0	6.02
HCT	%	±2.0	18.2	18.2	20.2
MCV	fL	±9.0	78.3	78.3	89.6
MCH	pg	±2.4	24.9	25.8	26.6
MCHC	g/dL	±3.0	31.9	33.0	29.8
PLT	10 ³ /μL	±30	62	62	70.5
MPV	fL	±3.0	6.3	8.6	8.3
RDW-CV *	%	±3.0	17.4	17.4	15.5
LY%	%	±10.0	60.6	60.6	60.90
MO%	%	±5.0	6.9	6.9	3.09
GR%	%	±10.0	32.5	32.5	36.01
LY	10 ³ /μL	±0.5	1.3	1.3	1.24
MO	10 ³ /μL	±0.2	0.1	0.1	0.06
GR	10 ³ /μL	±0.5	0.7	0.7	0.74

MEK-1305
QC ESR
50±17 mm

Manual Registration	
Lot	258
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



0694-901684U