

MK-RE2
HEMATOLOGY CONTROL
 for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE25Y-2
	2026-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	4.30	±2.70
RET	10⁶/μL	0.1497	±0.0940

2D Barcode Entry


• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
 It may have deteriorated or hemolyzed, which affects the measurement accuracy.


<https://www.nihonkohden.com/spi/index.html>
Manual Registration

Lot	25Y					
Level	2					
Expiration	1	M	5	D	2026	Y
Control	MKRE					


 0123



NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN

MK-RE3

HEMATOLOGY CONTROL
for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE25Y-3
	2026-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	8.64	±3.36
RET	10 ⁶ /μL	0.3026	±0.1177

2D Barcode Entry



• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

Manual Registration

Lot	25Y					
Level	3					
Expiration	1	M	5	D	2026	Y
Control	MKRE					

CE
0123

IVD



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

MK-RE1

HEMATOLOGY CONTROL
for reticulocyte

3.0mL

ASSAY SHEET

LOT	MKRE25Y-1
	2026-01-05
OPEN	14 DAYS

		MEK-9200	RANGE
RET%	%	1.61	±1.39
RET	10⁶/μL	0.0568	±0.0491

2D Barcode Entry



• If RBC value is under 3.00 [$\times 10^6/\mu\text{L}$], please refrain from using.
It may have deteriorated or hemolyzed, which affects the measurement accuracy.



<https://www.nihonkohden.com/spi/index.html>

Manual Registration

Lot	25Y					
Level	1					
Expiration	1	M	5	D	2026	Y
Control	MKRE					

CE
0123

IVD



NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN