

MEK-3DN HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B262N
	2026-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.8	7.6	7.6	7.61
RBC	10 ⁶ /μL	±0.20	4.80	4.80	4.75
HGB	g/dL	±0.5	13.6	13.7	13.85
HCT	%	±3.0	41.0	41.0	46.3
MCV	fL	±7.0	85.5	85.5	97.5
MCH	pg	±2.4	28.3	28.5	29.2
MCHC	g/dL	±3.0	33.2	33.4	29.9
PLT	10 ³ /μL	±50	236	236	249.9
MPV	fL	±2.0	5.7	8.2	8.2
RDW-CV *	%	±3.0	16.7	16.7	17.0
LY%	%	±10.0	27.7	27.7	31.24
MO%	%	±5.0	2.6	2.6	2.50
GR%	%	±10.0	69.7	69.7	66.26
LY	10 ³ /μL	±1.0	2.1	2.1	2.38
MO	10 ³ /μL	±0.4	0.2	0.2	0.19
GR	10 ³ /μL	±1.0	5.3	5.3	5.04

MEK-1305
QC ESR
10±5 mm

Manual Registration	
Lot	262
Level	Normal
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN



CE
 0123 **IVD**
 0694-901685W

MEK-3DH	HEMATOLOGY CONTROL	2.0mL
----------------	---------------------------	--------------


ASSAY SHEET

LOT	B262H
	2026-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order.
 If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±2.2	20.5	20.5	20.29
RBC	10 ⁶ /μL	±0.25	5.88	5.88	5.91
HGB	g/dL	±0.7	18.7	18.7	19.13
HCT	%	±4.0	54.7	54.7	63.0
MCV	fL	±7.0	93.1	93.1	106.6
MCH	pg	±2.8	31.8	31.8	32.4
MCHC	g/dL	±3.0	34.2	34.2	30.4
PLT	10 ³ /μL	±70	573	573	589.8
MPV	fL	±3.0	5.9	8.3	8.5
RDW-CV *	%	±3.0	16.0	16.0	17.8
LY%	%	±10.0	15.2	15.2	15.63
MO%	%	±5.0	1.0	1.0	1.38
GR%	%	±10.0	83.8	83.8	82.99
LY	10 ³ /μL	±1.8	3.1	3.1	3.17
MO	10 ³ /μL	±0.7	0.2	0.2	0.28
GR	10 ³ /μL	±2.0	17.2	17.2	16.84

Manual Registration	
Lot	262
Level	High
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>


NIHON KOHDEN CORPORATION

 1-31-4 NISHIOCHIAI, SHINJUKU-KU
 TOKYO 161-8560 JAPAN


MEK-3DH



B262H



2026-05-05



0694-901686W

MEK-3DL HEMATOLOGY CONTROL 2.0mL

ASSAY SHEET

LOT	B262L
	2026-05-05
OPEN	14DAYS

MEK-1301,1302,1305 Barcode Entry

The barcodes may be read in any order. If the barcode is difficult to scan, place this assay sheet on a flat surface, keep the barcode reader and barcode vertical, and scan within a distance of 2cm.

1



2



3



		RANGE	MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -1301 -1302 -1305
WBC	10 ³ /μL	±0.6	2.2	2.2	2.27
RBC	10 ⁶ /μL	±0.15	2.37	2.37	2.31
HGB	g/dL	±0.4	5.8	6.0	6.05
HCT	%	±2.0	18.2	18.2	20.1
MCV	fL	±9.0	76.9	76.9	87.1
MCH	pg	±2.4	24.5	25.3	26.2
MCHC	g/dL	±3.0	31.9	33.0	30.1
PLT	10 ³ /μL	±30	65	65	74.5
MPV	fL	±3.0	5.8	8.3	8.1
RDW-CV *	%	±3.0	18.1	18.1	16.4
LY%	%	±10.0	55.1	55.1	59.99
MO%	%	±5.0	6.4	6.4	3.08
GR%	%	±10.0	38.5	38.5	36.93
LY	10 ³ /μL	±0.5	1.2	1.2	1.36
MO	10 ³ /μL	±0.2	0.1	0.1	0.07
GR	10 ³ /μL	±0.5	0.9	0.9	0.84

MEK-1305
QC ESR
49±17 mm

Manual Registration	
Lot	262
Level	Low
Control	MEK-3D

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450.



<https://www.nihonkohden.com/spi/index.html>



NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN



MEK-3DL



B262L



2026-05-05



0694-901684W