

MEK-CAL

HEMATOLOGY CALIBRATOR

2.0mL

ASSAY SHEET

LOT	PLUS261
	2026-02-05
OPEN	7DAYS

MEK-9100,9200
2D Barcode Entry

MEK-1301,1302,1303,1305
Barcode Entry

<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION

1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN


MEK-CAL



PLUS261



2026-02-05



0694-901832H

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 ³ /μL	8.9	8.9	8.9	8.8	9.08	8.88
RBC	10 ⁶ /μL	4.55	4.55	4.54	4.50	4.49	4.48
HGB	g/dL	13.0	13.2	13.0	12.9	13.31	13.12
HCT	%	39.1	39.1	39.4	38.3	44.0	43.9
MCV	fL	86.0	86.0	86.8	85.2	97.9	98.1
MCH	pg	28.6	29.0	28.6	28.7	29.6	29.3
MCHC	g/dL	33.2	33.8	33.0	33.7	30.3	29.9
PLT	10 ³ /μL	226	226	225	224	235.2	241.9
MPV	fL	6.1	8.5	7.3	6.1	8.5	8.6
RDW-CV *	%	16.4	16.4	16.6	15.5	16.5	16.5

MEK -7300 -8222 -9100 -9200	
FS	170
FL	85
SD	109
FS THR.	100

* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,8222.