

MEK-CAL HEMATOLOGY CALIBRATOR 2.0mL

ASSAY SHEET

LOT	PLUS265
	2026-06-05
OPEN	7DAYS

		MEK -6400 -6410 -6420 -6450	MEK -6500 -6510 -6550	MEK -7300	MEK -8222	MEK -9100 -9200	MEK -1301 -1302 -1303 -1305
WBC	10 ³ /μL	9.1	9.1	9.1	9.0	9.31	9.21
RBC	10 ⁶ /μL	4.69	4.69	4.63	4.61	4.63	4.68
HGB	g/dL	13.8	14.0	13.7	13.6	13.97	14.04
HCT	%	41.6	41.6	41.1	40.0	45.8	46.6
MCV	fL	88.7	88.7	88.8	86.7	99.0	99.5
MCH	pg	29.4	29.9	29.6	29.5	30.2	30.0
MCHC	g/dL	33.2	33.7	33.3	34.0	30.5	30.1
PLT	10 ³ /μL	244	244	234	233	245.2	255.1
MPV	fL	6.2	8.5	7.3	6.2	8.6	8.6
RDW-CV*	%	16.6	16.6	16.6	15.5	16.5	16.4

MEK	
	-7300
	-8222
	-9100
	-9200
FS	168
FL	86
SD	106
FS THR.	100

MEK-9100,9200
2D Barcode Entry



MEK-1301,1302,1303,1305
Barcode Entry



* RDW-CV is displayed as RDW on MEK-6400,6410,6420,6450,8222.



<https://www.nihonkohden.com/spi/index.html>

NIHON KOHDEN CORPORATION
1-31-4 NISHIOCHIAI, SHINJUKU-KU
TOKYO 161-8560 JAPAN

